



Arginine Decarboxylase Test

Rapid test for detecting arginine decarboxylase activity.

INTENDED PURPOSE

Rapid test for evidence arginine decarboxylase production on bacterial isolates. This test is intended as an aid in the diagnosis, requiring further tests to complete the diagnostic results.

DESCRIPTION

Arginine Decarboxylase Test is used for the identification and differentiation of Enterobacteriaceae and other gram-negative organisms by means of decarboxylase reactions.

KIT CONTENT

- 30 Tubes of Arginine Decarboxylase Test

METHOD PRINCIPLE

The enzyme arginine decarboxylase attacks the carboxylic group of the amino acid arginine with formation (combination with dihydrolase) of amine putrescine. The sugar contained in the medium is fermented by all the enterobacteria with consequent initial colour change of indicator system from purple to yellow. The acid medium supports the amino acid decarboxylation reactions with formation of putrescine. The amines production alkalizes the medium and induces a new indicator colour change from yellow to purple. The negative reaction is shown by the appearance of a yellow colour in the tube, while the positive reaction is shown by a purple colour.

MATERIALS REQUIRED BUT NOT PROVIDED

Standard microbiological supplies and equipment such as: inoculating loop, pipettes, culture media, physiological solution, mineral (vaseline) oil, quality control organisms.

REAGENTS

- A tube of Lysine Decarboxylase Test contains desiccated biochemical substrates and nutritive components including arginine, glucose, bromocresol purple and pyridoxal.

SPECIMEN

Collect specimens in sterile containers or with sterile swabs and transport to the laboratory. Process each specimen using procedures appropriate for that sample. This product is recommended for use only with pure cultures.

Refer to specific guidelines for more detailed information.

TEST PROCEDURE

1. Take the number of Arginine Decarboxylase Test tubes needed from the fridge and allow them to reach room temperature.
2. Add 0.3 mL of physiological solution to each tube.
3. Inoculate with the test organism from a fresh overnight pure culture.
4. Overlay each inoculated tube with (2-3 drops) sterile mineral oil to separate the medium from external oxygen and promote fermentation.
5. Incubate tubes with caps tightened at $35 \pm 2^\circ\text{C}$ for up to 96 hours.
6. Examine tubes after 4 hours incubation and then daily for colour reactions.

INTERPRETING RESULTS

A positive result is indicated by the development of a purple colour.

A negative test is given by the development of a yellow colour for glucose-fermenting microorganisms, while non-glucose-fermenters will result in little or no colour change as compared to an uninoculated tube.

STORAGE

2-8°C in its original packaging. Keep away from sources of heat and avoid excessive changes of temperature. Use until the expiry date indicated on the label. Eliminate without using if there are signs of deterioration.

SHELF LIFE

1 year.

QUALITY CONTROL

Appearance: Purplish powder, yellowish-red solution once reconstituted

Control strain		Incubation	Characteristic reactions
<i>Enterobacter cloacae</i>	ATCC® 13047	18-24 h/ 35 ± 2°C	Positive reaction: Development of a purple color
<i>Pseudomonas aeruginosa</i>	ATCC® 27853		
<i>Proteus mirabilis</i>	ATCC® 25923		Negative reaction: Development of a yellow color
<i>Klebsiella pneumoniae</i>	ATCC® 13383		

PERFORMANCE CHARACTERISTICS

Performance testing of Arginine Decarboxylase Test was carried out using the QC strains listed above. The results obtained met the established criteria.

LIMITATIONS

Use of a heavy inoculum may be necessary to obtain a positive result in 4 h.

Interpretation earlier than 12-18 h incubation may lead to false-negative results for organisms that ferment glucose.

Exposure to air may cause alkalization of the surface of the medium giving false-positive results.

WARNING AND PRECAUTIONS

- 1) **For *in vitro* diagnostic use (IVD).**
- 2) **For laboratory professional use only.**
- 3) Operators must be trained and have certain experience. Please read the instructions carefully before using the product. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this document.
- 4) Consult the Safety Data Sheet (SDS) for information regarding hazards and safe handling practices.
- 5) Do not use if the product or packaging appears to be damaged.
- 6) Follow standard precautions. All patient specimens should be considered potentially infectious and handled accordingly.
- 7) Handle all specimens as if infectious using safe laboratory procedures. Dispose of hazardous or biologically contaminated materials according to the practices of your institution.
- 8) Avoid cross-contamination of samples by using disposable tips and changing them after each sample.
- 9) Do not mix reagents of different batches. Please use the product within the validity period.
- 10) Do not eat, drink, smoke, apply cosmetics or handle contact lenses in areas where reagents and human specimens are handled.
- 11) Results should be interpreted by a trained professional in conjunction with the patient's history and clinical signs and symptoms, and epidemiological risk factors.
- 12) Ensure laboratory equipment is calibrated and maintained in accordance with the laboratory's procedure.
- 13) When test results are transmitted from the laboratory to an informatics centre, attention has to be done to avoid erroneous data transfer.

DISPOSAL OF WASTE

Disposal of waste must be carried out according to national and local regulations in force.

BIBLIOGRAPHY

See the references at the end of this document.

TABLE OF SYMBOLS

See the table of symbols at the end of this document.

ORDER INFORMATION

Product	Packaging	Ref.
Arginine Decarboxylase Test	30 tests	88016

Revision History

Revision	Release Date	Change Summary
0	2023-05-15	Updated layout and content in compliance with IVDR 2017/746, version reset to revision 0

This IFU document and the SDS are available from the online Support Center:

liofilchem.com/ifu-sds



Istruzioni per l'uso
ITALIANO

Arginine Decarboxylase Test

Test rapido per la rilevazione dell'attività dell'arginina decarbossilasi.

DESTINAZIONE D'USO

Test rapido per evidenziare la produzione di arginina decarbossilasi su isolati batterici. Il test è inteso come ausilio alla diagnosi, e sono necessari ulteriori test per completare i risultati diagnostici.

DESCRIZIONE

Il test Arginine Decarboxylase è utilizzato per l'identificazione e la differenziazione delle Enterobacteriaceae e di altri organismi Gram-negativi mediante reazioni di decarbossilasi.

CONTENUTO DEL KIT

- 30 provette di Arginine Decarboxylase Test

PRINCIPIO DEL METODO

L'enzima arginina decarbossilasi attacca il gruppo carbossilico dell'aminoacido arginina con formazione (combinazione con diidrolasi) dell'ammina putrescina. Lo zucchero contenuto nel terreno viene fermentato da tutti gli enterobatteri con conseguente iniziale viraggio del sistema indicatore dal viola al giallo. Il terreno acido supporta le reazioni di decarbossilazione degli amminoacidi con formazione di putrescina. La produzione di ammine alcalinizza il terreno e induce un nuovo viraggio del colore dell'indicatore dal giallo al viola. La reazione negativa è evidenziata dalla comparsa di un colore giallo nella provetta, mentre la reazione positiva è evidenziata da un colore viola.

MATERIALI RICHIESTI MA NON FORNITI

Forniture e attrezzature microbiologiche standard quali: ansa per inoculo, pipette, terreni di coltura, soluzione fisiologica, olio minerale (vaselina), organismi per il controllo di qualità.

REAGENTI

- Una provetta di Lysine Decarboxylase Test contiene substrati biochimici essiccati e componenti nutritivi tra cui arginina, glucosio, porpora di bromocresolo e piridossale.

CAMPIONI

Raccogliere i campioni prelevati in contenitori sterili o con tamponi sterili e trasportarli in laboratorio. Trattare i campioni secondo la procedura appropriata per ognuno di loro. Questo prodotto è consigliato per l'uso esclusivamente con colture pure.
Fare riferimento alle linee guida specifiche per informazioni più dettagliate.

PROCEDURA DEL TEST

1. Prelevare dal frigorifero il numero necessario di provette per il test di arginina decarbossilasi e lasciare che raggiungano la temperatura ambiente.
2. Aggiungere 0,3 mL di soluzione fisiologica a ciascuna provetta.
3. Inoculare con il microrganismo da testare da una coltura pura fresca durante la notte.
4. Ricoprire ciascuna provetta inoculata con (2-3 gocce) di olio minerale sterile per separare il terreno dall'ossigeno esterno e promuovere la fermentazione.
5. Incubare le provette con i tappi serrati a $35 \pm 2^\circ\text{C}$ per un massimo di 96 ore.
6. Esaminare le provette dopo 4 ore di incubazione e poi giornalmente per le reazioni cromatiche.

INTERPRETAZIONE DEI RISULTATI

Un risultato positivo è indicato dallo sviluppo di un colore viola.

Un test negativo è dato dallo sviluppo di un colore giallo per i microrganismi che fermentano il glucosio, mentre i microrganismi che non fermentano il glucosio daranno luogo a un cambiamento di colore minimo o nullo rispetto ad una provetta non inoculata.

CONSERVAZIONE

Conservare a $2-8^\circ\text{C}$ nella sua confezione originale. Tenere lontano da fonti di calore ed evitare eccessivi sbalzi di temperatura. Utilizzare fino alla data di scadenza indicata sull'etichetta. Eliminare senza utilizzare se vi sono segni di deterioramento.

VALIDITÀ

1 anno.

CONTROLLO QUALITÀ

Aspetto: polvere violacea, soluzione rosso-giallastra una volta ricostituita

Ceppi di controllo		Incubazione	Caratteristiche di reazione
<i>Enterobacter cloacae</i>	ATCC® 13047	18-24 h/ $35 \pm 2^\circ\text{C}$	Reazione positiva: Sviluppo del colore viola
<i>Pseudomonas aeruginosa</i>	ATCC® 27853		
<i>Proteus mirabilis</i>	ATCC® 25923		Reazione negativa: Sviluppo del colore giallo
<i>Klebsiella pneumoniae</i>	ATCC® 13383		

CARATTERISTICHE PRESTAZIONALI

Il controllo delle prestazioni dell'Arginine Decarboxylase Test è stato eseguito utilizzando i ceppi sopra elencati. I risultati ottenuti hanno soddisfatto i criteri stabiliti.

LIMITAZIONI

Può essere necessario utilizzare un inoculo pesante per ottenere un risultato positivo in 4 h.

L'interpretazione prima dell'incubazione di 12-18 h può portare a risultati falsi negativi per i microrganismi che fermentano il glucosio.

L'esposizione all'aria può causare l'alcalinizzazione della superficie del terreno dando risultati falsi positivi.

AVVERTENZE E PRECAUZIONI

- 1) **Per uso diagnostico in vitro (IVD).**
- 2) **Solo per uso professionale di laboratorio.**
- 3) Gli operatori devono essere formati e avere una certa esperienza. Si prega di leggere attentamente le istruzioni prima di utilizzare il prodotto. L'affidabilità dei risultati del test non può essere garantita in caso di deviazioni dalle istruzioni contenute in questo documento.
- 4) Consultare la scheda di sicurezza (SDS) per informazioni sui pericoli e sulle pratiche di manipolazione sicure.
- 5) Non utilizzare se il prodotto o la confezione sembrano danneggiati.

- 6) Seguire le precauzioni standard. Tutti i campioni dei pazienti devono essere considerati potenzialmente infetti e maneggiati di conseguenza.
- 7) Maneggiare tutti i campioni come infetti utilizzando procedure di laboratorio sicure. Smaltire materiali pericolosi o biologicamente contaminati secondo le pratiche del proprio istituto.
- 8) Evitare la contaminazione incrociata dei campioni utilizzando puntali monouso e sostituendole dopo ogni campione.
- 9) Non mescolare reagenti di lotti diversi. Si prega di utilizzare il prodotto entro il periodo di validità.
- 10) Non mangiare, bere, fumare, applicare cosmetici o maneggiare lenti a contatto nelle aree in cui vengono manipolati reagenti e campioni umani.
- 11) I risultati devono essere interpretati da un professionista qualificato insieme alla storia del paziente, ai segni e sintomi clinici e ai fattori di rischio epidemiologici.
- 12) Assicurarci che le apparecchiature di laboratorio siano calibrate e mantenute in conformità con la procedura del laboratorio.
- 13) Quando i risultati dei test vengono trasmessi dal laboratorio a un centro informatico, è necessario prestare attenzione per evitare trasferimenti di dati errati.

SMALTIMENTO DEI RIFIUTI

Lo smaltimento dei rifiuti deve essere effettuato in conformità alle normative nazionali e locali in vigore.

INFORMAZIONI PER L'ORDINE

Prodotto	Confezionamento	Rif.
Arginine Decarboxylase Test	30 tests	88016












Questo documento IFU e la SDS sono disponibili dal Support Center online:

liofilchem.com/ifu-sds

References / Riferimenti

1. Lal A, Cheeptham N (2015) Decarboxylase Broth Protocol. American Society for Microbiology.
2. Lenette EH (1995) Manual of Clinical Microbiology.
3. Blazevic DJ, Ederer GM (1975) Principles of biochemical tests in diagnostic microbiology. 63-67. New York, John Wiley & Sons.
4. Brooks K, Sodeman T (1974) A rapid method for determining decarboxylase and dihydrolase activity. J Clin Pathol. 27(2):148-52. DOI: 10.1136/jcp.27.2.148.
5. Goldschmidt MC, Lockhart BM (1971) Rapid methods for determining decarboxylase activity: arginine decarboxylase. Appl Microbiol. 22(3):350-7. DOI: 10.1128/am.22.3.350-357.1971.

Table of Symbols / Tabella dei Simboli

	Batch code / Codice del lotto
	Catalogue number / Numero di catalogo
	<i>In Vitro</i> Diagnostic Medical Device / Dispositivo Medico Diagnostico <i>in vitro</i>
	Manufacturer / Fabbricante
	Use by / Utilizzare entro
	Fragile, handle with care / Fragile, maneggiare con cura
	Temperature limitation / Limiti di temperatura
	Contains sufficient for <n> tests / Contenuto sufficiente per <n> saggi
	Consult instructions for use / Consultare le istruzioni per l'uso
	Do not reuse / Non riutilizzare
	Keep away from sunlight / Tenere al riparo dalla luce solare



Liofilchem® s.r.l.

Via Scozia, 64026 Roseto degli Abruzzi (TE) Italy
Tel. +39 0858930745 Fax +39 0858930330

www.liofilchem.com

liofilchem@liofilchem.com

